



## AUTOMATED EXTERNAL DEFIBRILLATOR (AED) MODEL FOR MICHIGAN HIGH SCHOOLS

### A. Minimum AED Requirements -

- (1.) Inside the school building - A sufficient number of AEDs to enable a member of the school staff or other person to obtain the AED and deliver it to any location in the school building within 1 minute of being notified of an apparent cardiac emergency.
- (2.) Outside the school building on school grounds / athletic fields - A sufficient number of AEDs in the possession of the athletic trainer, coach, or other qualified person to enable someone to deliver the AED to any location outside of the school on school grounds including any athletic field, within 1 minute of being notified of an apparent cardiac emergency.
- (3.) Back-up AEDs - One or more AEDs to be held in reserve for use as a replacement for any AED which may be out of service due to maintenance or other issues. The back-up AED should also be made available for use by the school's athletic teams for traveling to off-site locations that may not have adequate AED coverage.

### B. Minimum AED Drill Requirements -

- (1.) At least one AED Drill per semester inside the school building with some student participation per semester.
- (2.) At least one AED Drill per semester outside the school building on school property such as on an athletic field with some student participation per semester.
- (3.) For information on how to conduct an AED Drill please see [www.aedrill.com](http://www.aedrill.com) or a similar website to be sponsored by the Michigan Department of Community Health.

